

## 205.10 - Radon Emanation (encapsulated solution form)

These SRMs are intended for the calibration of radon-222 measuring instruments.

[Click here](#) for more information.

Technical Contact: [ronald.colle@nist.gov](mailto:ronald.colle@nist.gov)

PLEASE NOTE: The tables are presented to facilitate comparisons among a family of materials to help customers select the best SRM for their needs. For specific values and uncertainties, the certificate is the only official source.

SRM	Description	Unit of Issue	Radium 226 Activity (Bq)	Time of Calibration (month/year)
4971	Radon-222 Emanation Standard	0.2 mL	5.192 ± 0.062	9/03
4972	Radon-222 Emanation Standard	0.2 mL	49.55 ± 0.59	9/03
4973	Radon-222 Emanation Standard	0.2 mL	517.7 ± 6.2	9/03
4974	Radon-222 Emanation Standard	0.2 mL	5761 ± 75	5/05